



Product Analysis Certificate

Art.Nr.: CL03.0204

Lot Nr.: 27.4741912

Buffer pH 10.00

pH 10.00 ($\pm 0.02/20^{\circ}\text{C}$) H₃BO₃-KCl-NaOH

Measured value: pH = 10,00 - 20°C

Use: For laboratory use, pH measurements, traceable to SRM from NIST

Traceability: Reference standards acc. NIST SRM 189c, 188, 185i, 186lg, 186llg, 187e

Specifications:

pH = 10.24 @ 0°C
pH = 10.10 @ 10°C
pH = 10.00 @ 20°C
pH = 9.95 @ 25°C
pH = 9.90 @ 30°C
pH = 9.82 @ 40°C
pH = 9.75 @ 50°C
pH = 9.68 @ 60°C
pH = 9.62 @ 70°C
pH = 9.55 @ 80°C
pH = 9.49 @ 90°C

Chemist: Luis Bianchi

Date of release: 07 January 2020

Expires: Dec-2021

Version: 1

CHEM-LAB NV

Industriezone "De Arend" 2 B-8210 ZEDELGEM - BELGIUM

Tel.: +32 50 28 83 20 Fax.: +32 50 78 26 54 e-mail : info@chem-lab.be Web : www.chem-lab.be